



**IR-750 Instruction Manual  
Supplement**

Title: IR-750 Instruction Manual  
 Part Number: 4279850A  
 Print Date: October, 2011  
 Revision/Date: 09-Jan-2014

Supplement Issue: 1  
 Issue Date: 01/2014  
 Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

**Change # 1**

On page # 27, under “**SPECIFICATION**” replace

**From:**

<p><b>Accuracy (assumes ambient operating temperature of 21°C to 25°C (70°F to 77°F))</b></p>	<p>&gt;0°C to 1550°C (&gt;32°F to 2822°F): ±1.8% or ±1.8°C (±4°F), whichever is greater                  &gt;-35°C to 0°C (-31°F to 32°F): ±1.8%+1°C (2°F) or ±2.8°C (±6°F), whichever is greater                  -50°C to -35°C (-58°F to -31°F): not specified (for reference only)</p>
---	--

**To:**

<p><b>Accuracy (assumes ambient operating temperature of 21°C to 25°C (70°F to 77°F))</b></p>	<p>&gt;0°C to 1550°C (&gt;32°F to 2822°F): ±1.8% or ±1.8°C (±4°F), whichever is greater (<b>Typical</b>)                  &gt;-35°C to 0°C (-31°F to 32°F): ±1.8%+1°C (2°F) or ±2.8°C (±6°F), whichever is greater (<b>Typical</b>)                  -50°C to -35°C (-58°F to -31°F): not specified (for reference only)</p>
---	--